


Code 2963LR-G

Total Chlorine
25 Test Strips



Insta-Test[®] ANALYTIC



Using a cup-size sample,
dip strip for **2 seconds**.

Remove strip with pad
face up. DO NOT SHAKE
OFF EXCESS WATER.

With pad still facing up,
wait **10 seconds** and read
the strip. Discard strip
after 30 seconds.



- Keep wet fingers out of the bottle.
- Cap bottle tightly.
- Store in a cool, dry place.

LaMotte Company
Chestertown, Maryland 21620 USA
www.lamotte.com

