



BROMTHYMOL BLUE pH KIT

OCTA-SLIDE 2, pH 6.0-7.4


CODE 2109-01

QUANTITY	CONTENTS	CODE
30 mL	Bromthymol Blue Indicator	2210-G
2	Test Tubes, 2.5-5-10 mL, plastic, w/caps	0106
1	Bromthymol Blue Octa-Slide 2 Bar, 6.0-7.4	2184-01
1	Octa-Slide 2 Viewer	1101

*WARNING: Reagents marked with an * are considered to be potential health hazards. To view or print a Material Safety Data Sheet (MSDS) for these reagents go to www.lamotte.com. To obtain a printed copy, contact LaMotte by e-mail, phone or fax.

To order individual reagents or test kit components, use the specified code number.







USE OF THE OCTA-SLIDE 2 VIEWER



The Octa-Slide 2 Viewer should be held so non-direct light enters through the back of the Viewer. Insert the reacted sample into the top of the Viewer. Slide the Octa-Slide 2 Bar into the Viewer and match the color of the reaction to the color standards.

WARNING! This set contains chemicals that may be harmful if misused. Read cautions on individual containers carefully. Not to be used by children except under adult supervision

PROCEDURE

<p>1.</p>  <p>Fill a test tube (0106) to the 10 mL line with sample water.</p>	<p>2.</p>  <p>Add 8 drops of Bromthymol Blue Indicator (2210)</p>
<p>3.</p>  <p>Cap and mix.</p>	<p>4.</p>  <p>Insert Bromthymol Blue Octa-Slide 2 Bar (2184-01) into the Octa-Slide 2 Viewer (1101)</p>
<p>5.</p>  <p>Insert test tube into Octa-Slide 2 Viewer (1101).</p>	<p>6.</p>  <p>Match sample color to a color standard. Record as pH.</p>

LaMOTTE COMPANY

Helping People Solve Analytical Challenges®

PO Box 329 • Chestertown • Maryland • 21620 • USA
800-344-3100 • 410-778-3100 (Outside U.S.A.) • Fax: 410-778-6394
Visit us on the web at www.lamotte.com