



HYDROGEN PEROXIDE KIT

DROP COUNT, 1 drop = 5 ppm

CODE 7138-DB-01

QUANTITY	CONTENTS	CODE
10 g	*Potassium Iodide Crystals	*6809-D
30 mL	Starch Indicator Solution	4170WT-G
60 mL	Peroxide Titrant	7139-H
60 mL	*Acidified Catalyst	*7140-H
1	Test Tube, 5-10-15-20-25 mL, plastic, w/cap	0715
1	Spoon, 0.1g, plastic	0699

*WARNING: Reagents marked with an * are considered to be potential health hazards. To view or print a Material Safety Data Sheet (MSDS) for these reagents go to www.lamotte.com. To obtain a printed copy, contact LaMotte by e-mail, phone or fax.

To order individual reagents or test kit components, use the specified code number.

PROCEDURE

1. Rinse and fill the test tube (0715) to 25 mL line with sample solution.
2. Add 10 drops of *Acidified Catalyst (7140). Swirl to mix.
3. Use the 0.1g spoon to add one level spoonful of *Potassium Iodide Crystals (6809). Swirl until crystals dissolve. Solution will turn a yellow to brown color if hydrogen peroxide is present.
4. Add 5 drops of Starch Indicator Solution (4170WT). Swirl to mix. Solution will turn light to dark blue.
5. While gently swirling the tube, add Peroxide Titrant (7139) one drop at a time, until blue color disappears. Count the number of drops added. Hold bottle vertically.
6. Multiply the number of drops used in Step 5 by 5. Record as ppm Hydrogen Peroxide (H_2O_2).

LaMOTTE COMPANY

Helping People Solve Analytical Challenges®

PO Box 329 • Chestertown • Maryland • 21620 • USA
800-344-3100 • 410-778-3100 (Outside U.S.A.) • Fax 410-778-6394

Visit us on the web at www.lamotte.com